

## **MEASUREMENT REPORT**

### **TRANSPORTATION**

#### **Conditions of the transfer standards on arrival:**

Date of arrival:	
Method of transportation:	
Transportation box:	
Other remarks:	

#### **Initial inspection of transfer standards:**

Weight	200 g	20 g	2 g	200 mg
Top				
Side envelope				/
Bottom				
Comments				

#### **Conditions of the transfer standards on departure:**

Date of departure:	
Method of transportation:	
Transportation box:	
Other remarks:	

#### **Final inspection of transfer standards:**

Weight	200 g	20 g	2 g	200 mg
Top				
Side envelope				/
Bottom				
Comments				

## CALIBRATION RESULTS

Nominal value	Mass (true) $m$	Standard uncertainty $u_m (k=1)$	Conventional mass $m_c$	Standard uncertainty $u_{m_c} (k=1)$	No. of measurement cycles *
200 g					
20 g					
2 g					
200 mg					

\*) If the calibration was made by other procedures than one-to-one comparison (e.g. subdivision using a system of weighing equations), give the degrees of freedom of this procedure.

## DATA PROVIDED BY MIRS

Weight	Volume $V$ (cm <sup>3</sup> )	Uncertainty $u_V$ (cm <sup>3</sup> ) ( $k = 1$ )	Centre of gravity $z$ (mm)	Uncertainty $u_z$ (mm) ( $k = 1$ )	Magnetic susceptibility $\chi$
200 g	25,102	0,006	21	2,0	$\chi < 0,02$
20 g	2,5091	0,0016	9,7	2,0	$\chi < 0,02$
2 g	0,2514	0,0009	4,6	1,0	$\chi < 0,06$
Weight	Density $\rho$ (kgm <sup>-3</sup> )	Relative uncertainty $\rho$ ( $k = 1$ )			
200 mg	8600	2 %			

The assumed volume temperature expansion coefficient for the 2 g, 20 g and 200 g weight is  $\alpha = 48 \pm 5 \cdot 10^{-6} \text{ K}^{-1}$

## TRACEABILITY OF REFERENCE STANDARDS

Insert the standard(s) including additional weights used for the calibration and its traceability to the international prototype of the kilogram

### 200 g

Identification	Mass $m$	Standard uncertainty $u_m (k=1)$	Volume $V$	Standard uncertainty $u_v (k=1)$	Calibration date

### 20 g

Identification	Mass $m$	Standard uncertainty $u_m (k=1)$	Volume $V$	Standard uncertainty $u_v (k=1)$	Calibration date

### 2 g

Identification	Mass $m$	Standard uncertainty $u_m (k=1)$	Volume $V$	Standard uncertainty $u_v (k=1)$	Calibration date

### 200 mg

Identification	Mass $m$	Standard uncertainty $u_m (k=1)$	Volume $V$	Standard uncertainty $u_v (k=1)$	Calibration date

## DESCRIPTION OF THE MEASUREMENT PROCEDURE

--

## UNCERTAINTY BUDGET

The uncertainties shall be estimated and combined following the Guide to the Expression of Uncertainty in Measurement JCGM 100:2008.

	(Standard) uncertainty contributions (mg)			
Uncertainty component	200 g	20 g	2 g	200 mg
Combined uncertainty $u_c (k=1)$				

## DETAILS OF THE BALANCE USED FOR THE CALIBRATION

Weight	Manufacturer	Type	Maximum load	Resolution	Standard deviation*	Manual/automatic
200 g						
20 g						
2 g						
200 mg						

\*Indicate the weighing procedure and number of measurements

## DETAILS OF THE INSTRUMENTS USED FOR AIR DENSITY DETERMINATION

	Manufacturer	Type	Range	Resolution	Standard uncertainty ( $k=1$ )
Temperature					
Pressure					
Humidity					
CO <sub>2</sub>					

## DESCRIPTION OF AIR BUOYANCY DETERMINATION

--

## ENVIRONMENTAL DATA DURING CALIBRATION

Insert the maximum and minimum values of the measured quantities and their standard uncertainties.

Temperature $t$ (°C)	Pressure $p$ (mbar)	Relative humidity $h$ (%)	$x(\text{CO}_2)$ $\times 10^6$	Air density $\rho_a$ (kg/m <sup>3</sup> )

## DESCRIPTION OF THE MEASURING ROOM

Mean temperature	
Minimum, maximum temperature	
Maximum change of temperature during 8 hours	
Mean humidity	
Minimum, maximum humidity	

Laboratory:	Responsible person:
Date:	Signature: