

Euromet Project No. 419

Final Report

Contents

1.	Introduction	2
2.	Presentation af the intercomparison.....	2
2.1.	<i>The participating laboratories.....</i>	2
2.2.	<i>The Calibration object.....</i>	3
2.3.	<i>Definitions</i>	3
2.4.	<i>The transfer of the standard</i>	4
3.	Calibration procedure	4
4.	Equipment.....	5
4.1.	<i>Description of the installation.....</i>	5
4.1.1.	FORCE Institutes, Denmark.....	5
4.1.2.	NMI, Netherlands.....	5
4.1.3.	Gaz De France, France.....	5
4.1.4.	CESAME-LNE QUEST, France	6
4.1.5.	Ministère des Affaires Economiques, Belgium.....	6
4.1.6.	Eidg. Amt für Messwesen, Switzerland.....	6
4.1.7.	Bundesamt für Eich und Vermessungswesen, Austria	6
4.1.8.	Instituto de Metrologia "G. Colonnelli", Italy.....	7
4.1.9.	Országos Mérésügyi Hivatal, Hungary	7
4.1.10.	Czech Metrological Institute, Czech Rep.....	7
4.1.11.	PTB, Germany.....	7
4.1.12.	Slovak Institute of Metrology, Slovakian.....	8
4.1.13.	Central Office of Measures, Poland.....	8
4.1.14.	NEL, United Kingdom.....	9
4.2.	<i>Uncertainty.....</i>	9
5.	Stability of the meter	9
The meter was calibrated three times at the FORCE-Institute		9
Calibration result.....		11
6.1	<i>Each laboratory</i>	11
6.1.1	Lab. 1. FORCE Institute, Denmark.....	11
6.1.2	Lab. 2. NMI.....	14
6.1.3	Lab. 3. Gaz de France, France.....	16
6.1.4	Lab. 4. CESAME-LNE QUEST, France	18
6.1.5	Lab. 5. Ministère des Affaires Economiques, Belgium.....	20
6.1.6	Lab. 6. Eidg. Amt für Messwesen, Switzerland.....	22
6.1.7	Lab. 7. Bundesamt für Eich und Vermessungswesen, Austria	24
6.1.8	Lab. 8. Instituto de Metrologia "G. Colonnelli", Italy.....	26
6.1.9	Lab. 9. Országos Mérésügyi Hivatal, Hungary	28
6.1.10	Lab. 10. Czech Metrological Institute, Czech Rep.	30
6.1.11	Lab. 11. PTB, Germany	32
6.1.12	Lab. 12. Slovak Institute of Metrology, Slovakian.....	34
6.1.13	Lab. 13. Central Office of Measures, Poland	36
6.1.14	Lab. 14. NEL, United Kingdom	38
6.1.15	All the laboratories	40
6.1.16	Average error, Median erro and weighing error.....	41

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

1. INTRODUCTION

A G250 Rotary Gas Meter shall be circulated to 14 laboratories in Europe for calibration. The laboratories use their usual calibration procedure, and shall, after the calibration produce a report containing a description of their installation and the calibration data for their measuring equipment. NMI has provided the meter, And FORCE-institute has prepared the proposal, and finished the final report.

2. PRESENTATION OF THE INTERCOMPARISON

2.1. The participating laboratories

Denmark:

FORCE Institute
Navervej 1
6600 Vejen

Italy:

Instituto di Metrologia "G. Colonnetti"
Consiglio Nazionale delle Ricerche
Strada delle Cacce 73
10135 Torino

Netherlands:

NMI
PO Box 394
3300 AJ Dordrecht

Hungary:

Országos Mérésügyi Hivatal
H-1535
Budapest Pf. 919

France:

Gaz De France
1. chemin de Villeneuve
94140 Alfortville

Czech Rep.:

Czech Metrological Institute
Prumyslova 455
530 03 Pardubice

France:

CESAME-LNE QUEST
43. route de l'aérodrome
86000 Poitiers

Germany:

PTB
Postfach 3345
38023 Braunschweig

Belgien:

Ministère des Affaires Economiques
Service de la Métrologie
Chaussée de Haecht 1795
1130 Bruxelles

Slovakian:

Slovak Institute of Metrology
Karloveska 63
84255 Bratislava

Switzerland:

Eidg. Amt für Messwesen
Lindenweg 50
3084 Wabern

Poland:

Central Office of Measures
Division of Thermodynamics
ul. Elektroralna 2
00950 Warszawa

Austria:

BEV
Arltgasse 35
Postfach 20
1163 Wien

United

Kingdom
NEL
East Kilbride
Glasgow G75 OQU

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

2.2. The Calibration object

Description of the meter G250 (Rotarymeter)

Manufacturer:	Instromet
Pipe size:	4"
Max Pressure:	10 bar
Capacity:	2-400 m ³ /h
Pulse generators:	1689,5 imp = 1 m ³ (2 · namur marked 1 and 2)
Pressure tapping:	Pr (inlet). - P (outlet)
Temperature measurement:	2 * d downstream the meter
Number:	001

2.3. Definitions

The way to calculate the error of the meter are:

$$E_m = \frac{\text{Indicated Volume} - \text{Real Volume}}{\text{Real volume}} \cdot 100\%$$

Indicated Volume are:

$$V_{im} = \frac{\text{Number of HF - Pulses}}{\text{Pulsevalue for the meter}^*} \text{ (m}^3\text{)}$$

* Pulsevalue 1689,5 pulse/1 m³

Real Volume:

Volume Calculated from the standards which are used in the laboratory.

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

2.4. The transfer of the standard

In the end of march 1998 FORCE Institutes did the first calibration of the meter.

All the laboratories was responsible for the transfer of the meter to the nest laboratory.

The meter was calibrated as listed below:

Laboratory	Day of calibration	Received	Send
NMI, Netherlands	97.07.03		
FORCE Institute, Denmark	98.03.23/27		98.05.12
Gaz de France, France	98.06.02/03	98.05.13	
CESAME-LNE OUEST, France	98.07.06	98.06.18	98.07.07
Ministère des Affaires Economiques, Belgien	98.	98.07.10	98.08.04
Eidg. Amt für Messwesen, Switzerland	98.08.20/21	98.08.10	98.08.26
BEV, Austria	98.09.01	98.08.28	98.09.17
Instiuto de Metrologia "G. Colonnelli", Italy	98.11.06	98.09.29	98.11.09
PTB, Germany	99.01.05	98.12.02	99.02.05
NEL, United Kingdom	99.03.09/31	99.02.18	99.04.12
Slovak Institute of Metrology, Slovakian		99.07.27	99.10.04
Czech Metrological Institute, Czech Rep.		99.10.27	99.12.06
Országos Mérésügyi Hivatal, Hungary	00.01.21	99.12.10	00.01.31
Central Office of Measures, Poland	00.03.06/07	00.02.05	

3. CALIBRATION PROCEDURE

The meter has to be calibrated in 11 points carried out 3 times.

The conditions is under atmospheric pressure and 20°C.

The flowrates is 2-5-10-20-40-80-120-160-240-320 and 400 m³/h.

The flowrate must be regulated within $\pm 2\%$, and the calibration shall be carried out starting with 400 m³/h then go down to 2 m³/h, go up to 400 m³/h again and then chose the flowrate by random.

Before the calibration starts, the meter shall be exercised minimum 1 hours at 240 m³/h.
The meter shall be calibrated without oil.

After the exercising a pressure drop test shall be made at 240 m³/h, the value shall be in the range 415-425 Pa.

Results of the pressure test shall be send to the projectmanager before the meter leave the laboratory.

The temperature must not vary more than 0.2°C under any calibration point, and the calibration temperature must be within 20°C $\pm 2\%$.

The pressure measuring must be at the Pr point.

For pulsregistration we use pulse generator namur: 1 m³ = 1689,5 pulses.
(Use the enclosed VSL-LD 3 and/or 4).

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

Under the calibration, following data shall be registered.

Barometric pressure in **Pa** or **mbar**.

Pressure (abs) at the Pr point in **PA** or **mbar**.

Pressure loss over the meter from Pr. to P in **Pa**.

Flow at the meter (indicated) in **m³/h**.

Temperature at the meter (2* d downstream) in **°C**.

Error at the meter: (indicated volume - real volume)/real volume * 100%.

Uncertainty in accordance with Wecc 19.

Used calibration equipment with data.

Traceability for all measuring equipment.

Adresses and timeschedule for the participating laboratories: Enclosure 1.

4. EQUIPMENT

4.1. Description of the installation

4.1.1. FORCE Institutes, Denmark.

The meter was calibrated against working standards:

The standard meters run in parallel (on meter at the time) and the meter to which have to bee calibrated, run in series with the standard meters. After each standard meter, and the meter under test a thermistor measure the temperature.

A Difference pressure transmitter measure the pressure-difference between the meter under test and the relevant standard meter.

Under the calibration, high frequency pulses from the meter, and the relevant standard meter are registered by a microcomputer which also registered temperature and pulses.

Everything runs automatically.

The whole calibration system have been calibrated against our National Standards meter.

4.1.2. NMI, Netherlands

4.1.3. Gaz de France, France

Gas de Franc have calibrated the meter at Natural gas type H. The main components have an average concentration (% molaire)

N₂	CH₄	CO₂	C₂H₆	C₃H₈
1,849	87,584	1,454	6,428	1,900

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

Density = 0,640

$\rho_n = 0,8276 \text{ kg/m}^3 (n)$

- a low pressure = 40 mbar
- the pipe ($\varnothing 50 \text{ mm}$) length upstream CVM meter is = 1,60 m (32 DN)
- the laboratory is air conditioned and the temperature are regulated within $20^\circ\text{C} \pm 2^\circ\text{C}$. The temperature of the gas are $20,1^\circ\text{C}$, with a standard deviation of $\pm 0,1^\circ\text{C}$.

The equipment used for calibration are a system and sonic nozzles in parallel.

4.1.4. CESAME-LNE QUEST, France

A set of Venturi nozzle operating in sonic conditions is used for the determination of the standard mass-flowrate.

Compressed dry air stored in a 110 m^3 vessel under 200 bar is used as the test fluid.

The meter under test is placed on a pipeline downstream the nozzles.

The pressure and the temperature are measured at the level of the meter in test in order to determine the volumetric flowrate.

4.1.5. Ministère des Affaires Economiques, Belgium

On the graduated scale of the bell are fixed reflecting flags separated each other by a length corresponding to a volume of the bell of 1 m^3 . When the bell moves downwards, these flags passed in front of a photoelectric cell and electric pulses are generated. The obtained pulses are used to start and stop the meter pulse counters. It is possible to program the volume of air used up the meter and the volume used for the test of the meter. The measurements of the temperature and the pressure of the bell are used to calculate the mass of air flowing out of the bell prover. The real volume of air flowing through the meter is calculated from the mass of air and the measurements of temperature and pressure of the meter. The pulses from the meter and their frequency give the indicated volume and the indicated meter flowrate.

Just after the connection of the meter and its pipes to the test rig, a leak test is realized.

4.1.6. Eidg. Amt für Messwesen, Switzerland

A 10m^3 bell prover blows the air through the meter under test. The bell prover is calibrated with 1000 litre oil filled volume standard. The immersion of the bell prover is measured with an electro-optic ruler (Heidenhain). The gage pressure of the air in the prover is 20 mbar. The temperatures of the air leaving the bell prover and of the air at the meter under test, the gage pressure of the bell prover, the pressure drop from the bell prover to the meter under test and the barometric pressure are registered every second. The pulses of the meter under test are counted with two counters. The totalizing of these pulses of the meter meter depends on a signal given from the electro-optic ruler of the bell prover (start-stop). A HP 310-computer calculates the volume at the meter under test by taking into account the mean values of the temperature and pressure differences, the influences of the drift of the barometric pressure, and the leak rate. Finally, the computer calculates the error of the meter under test and writes all the relevant data on a printer.

4.1.7. Bundesamt für Eich und Vermessungswesen, Austria

Equipment used for the calibration:

Flowrange $0,1 \text{ m}^3/\text{h}$ to $1000 \text{ m}^3/\text{h}$ under atmospheric conditions.

N1 = Turbine meter G650

N2 = Turbine meter G250

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

N3 = Rotary seal meter G65
N4 = Rotary meter G16
N5 = Drum meter NB3

4.1.8. Instituto de Metrologia "G. Colonnelli", Italy

The standard /1/ used by the IMGC is a prover, whose volumetric device is a 1000 mm diameter and 1630 mm long piston. This is forced by a brushless motor and a lead screw to sink downwards at a pre-set and controlled velocity into a measurement chamber, whence it displaces known volumes of gas in known time intervals. Actual operation is as follows.

First of all piston is raised up to the top end of its stroke, while filtered air is admitted into the measurement chamber. The pressure is atmospheric; a period of a few minutes is allowed for the temperature to stabilize. Then the admission valve is closed and a large output valve (100 mm nominal diameter) connected to the bottom of the cylindrical measurement chamber is opened. The gas meter under test is located downstream this output valve, being inserted in a 100 mm bore horizontal pipe; the straight lengths are 15 D upstream and 10 D downstream the meter. The pipe is externally coated with 3 cm thick neoprene.

Then, the digital electronic accelerates rapidly the motor up to the pre-set constant velocity, while pressure in the chamber builds up and air begins to flow through the gas meter. As soon as gas pressure, flowrate and gas revolution speed are stabilized to their steady-state values, measurements are started by gating a pre-set number of pulses (ranging between 400 and 900) from the gasmeter. In the same time interval, measured by a clock (and hereafter called useful measurement period), the pulses emitted by a rotating encoder fitted to the female screw of the piston drive are counted. Signals from four manometers and 14 temperatur transducers of the PRT type acquired are acquired as well at various times.

/1/ Cignolo, G., Rivetti, A., Martini, G., Alasia, F., Birello, G., La Piana, G. "The national standard gas provers of the IMGC-CNR", Flomeko 2000, Salvador (Brazil), June 2000

4.1.9. Országos Mérésügyi Hivatal, Hungary

OMH have calibrated the meter at OMH's Verification Laboratory situated at Budapest Gas Company. The test rig was set up with rotatorymeter upstream of the reference meters. Temperatur was measured upstream and 5D downstream of the meters. Pressure was measured at the p_r.

The kind of reference standards:

1. Turbine gas meter (60 - 500 m³/h)
2. Rotary gas meter (2,4 - 50 m³/h)

4.1.10. Czech Metrological Institute, Czech Rep.

The testing bench with sonic nozzles consists of 14 nozzles which are situated in 3 blocks. The vacuum is generated by two centrifugal fans and by one vacuum pump. The clamping system of gas meter is pneumatic. There are one barometric pressure meter and six gauge pressure sensors. Three of gauge pressure sensors measure the underpressure in blocks of nozzles, one of them measures the tightness of lines which are out of operation, one measures the underpressure in gas meter (p_r) and the last one measures the pressure loss of gas meter. Five temperature sensors measure the temperature in blocks of nozzles, in the gas meter and in the input of air to testing bench. Besides the humidity in the input of air to the testing bench and the time of test are measured, too.

4.1.11. PTB, Germany

The gas meter test rig with critical nozzles.

The critical nozzles for flowrate up to 100 m³/h are calibrated by direct installation into the primary standard bell prover of PTB without any additional steps, the larger nozzles by a precise step-by-step procedure using the small nozzles and very stable CVM gas meters. The uncertainty of the secundary standard test rig is 0,08% (expanded uncertainty with k=2 in accordance with WECC 19)

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

4.1.12. Slovak Institute of Metrology, Slovakian.

Volume flowrate: 0,02 to 20 m³/h
Manufacture: Justur, Stara Tura, Slovak republic
Test fluid: air under atmospheric conditions, is sucked from laboratory to test rig by vacuum pumps
Standards: 10 critical Venturi nozzles in parallel with individual flowrates between 0,02 to 7 m³/h
Operation: The equipment is portable box with pressure and teperature transmitters and vacuum pumps
Control and data acquistion: in-line by PC
Traceability for critical nozzles: to PTB's primary standards, Braunschweigh, Germany in March 1999.

Description of the equipment: Rotary meter G 100

The rotary gas meter, with rotary chambers (CVM system), size G 100

Volume flowrate: 16 to 160 m³/h
Manufacture: Elster AG Production, Mainz-Kastel, Germany
Pipe conection: DN 80
Manuf. Number: 19.970.709
Year of manuf.: 1998
Test fluid: air under atmospheric conditions
Traceability for meter: to PTB's primary standards, Braunschweigh, Germany in May 1998.

Description of the equipment: Turbine meter G 650

The axial turbine gas meter, size G 650

Volume flowrate: 100 to 1 000 m³/h
Manufacture: Elster AG Production, Mainz-Kastel, Germany
Pipe conection: DN 150
Manuf. Number: 83.026.952
Year of manuf.: 1998
Test fluid: air under atmospheric conditions
Traceability for meter: to PTB's primary standards, Braunschweigh, Germany in May 1998.

Place of calibration:

The rotary meter Instromet have been calibrated at metrological laboratory of Premagas, Stara Tura, Slovak republic, by SMÚ's standards (Sonic nozzles) and turbine and rotary standards of Premagas. From August 1999 the SMÚ's test rig for higher flowrates is not function, because in gas laboratory

of our institue was built chamber with new primary standards.

4.1.13. Central Office of Measures, Poland

The bell prover used for calibration is characterised by the following parameters:

- | | |
|---|------------------------------------|
| - Maximum measured volume | - 65 m ³ |
| - Minimum measured volume | - 5 m ³ |
| - Pressure at the bell | - 4 kPa |
| - The length of the scale | - 5 m |
| - Method of selecting volume | - electromechanical contacts |
| - Seal of the bell | - by water |
| - Flow range | - 9 to 7000 m ³ /h |
| - The stability of the gas pressure inside the bell | - by cone-shaped compensation pipe |
| - Operating mode | - manual |

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

4.1.14. NEL, United Kingdom

The test rig was set up with the positive displacement (PD) meter placed downstream of the reference sonic nozzles. Temperature and pressure were measured upstream of the reference meter. At the PD meter pressure was measured at the p_r, while temperature was measured 5D downstream of the meter. The differential pressure across the PD meter was measured between points Pr and P.

4.2. Uncertainty

The uncertainty of the different installation are put in the scheme as combined uncertainty.

Laboratory	Flow m ³ /h	Combined %
FORCE Institutes Denmark	2 – 10 10 – 400	0,17 0,13
NMI Netherlands	2 - 400	0,10?
Gaz de France France	10 20 - 400	0,38 0,34
CESAME-LNE OUEST France	10-400	0,22
Ministère des Affaires Economiques, Belgian	2 – 400	0,19??
Eidg. Amt für Messwesen Switzerland	2 – 5 10 – 400	0,18 – 0,13 0,10 – 0,11
BEV Austria	2 – 400	0,30
IMGC Italy	5 10 – 40 80 120 160	0,06 0,025 0,075 0,1
Országos Mérésügyi Hivatal Hungary	2 – 400	0,31
Czech Metrological Institute Czech Rep.	2 – 400	0,26
PTB Germany	2 – 400	0,08
Slovak Institute of metrology Slovakian	2 – 160 240 - 400	0,21 0,22
Central Office of Measures Poland	9 - 400	0,22 – 0,26
NEL United Kingdom	2 – 400	0,37

5. STABILITY OF THE METER

The meter was calibrated three times at the FORCE-Institute

First time	98.03.26/27
Sekund time	99.06.01/03
Third time	00.04.10

The stability is within 0,05% in the range from 5 – 400 m³/h.


Results: error in %

Flow (m ³ /h)	First time	Sekund time	Third time	Average
2	-0,41	-0,72	-0,76	-0,63
5	-0,06	-0,14	-0,06	-0,09

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

10	-0,04	-0,01	0,02	-0,01
20	0,00	-0,01	0,02	0,00
40	0,13	0,10	0,11	0,11
80	0,10	0,16	0,18	0,15
120	0,11	0,18	0,18	0,16
160	0,10	0,18	0,17	0,15
240	0,22	0,28	0,28	0,26
320	0,35	0,39	0,38	0,37
400	0,51	0,61	0,57	0,56



In this project only results from the first calibration is used in the comparison.

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

CALBRATION RESULT

6.1 *Each laboratory*

Calibration result for all the participating laboratories are listed on the following pages.

6.1.1 Lab. 1. FORCE Institute, Denmark

Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

	Flow real m³/h	Pressure abs. at Pr mbar	Pressure drop. Pr-P Pa	Temperature °C	Error %	Standard deviation %
max. - min.	2,5	1017,0	1,0	20,0	-0,50	
min. - max	2,2	1017,0	1,0	20,0	-0,39	
random	2,1	1016,1	1,0	20,0	-0,34	
Average	2,3	1016,7	1,0	20,0	-0,41	0,08
max. - min.	5,3	1017,0	1,0	20,0	-0,04	
min. - max	5,0	1017,0	1,0	20,0	0,02	
random	5,2	1016,1	3,0	20,0	-0,15	
Average	5,2	1016,7	1,7	20,0	-0,06	0,09
max. - min.	9,7	1017,0	2,0	20,0	-0,04	
min. - max	9,7	1017,0	2,0	20,0	-0,04	
random	9,2	1016,1	1,0	20,1	-0,05	
Average	9,5	1016,7	1,7	20,0	-0,04	0,01
max. - min.	19,9	1017,0	4,0	20,0	0,00	
min. - max	19,7	1017,0	4,0	20,0	0,00	
random	19,6	1016,1	4,0	20,0	0,01	
Average	19,7	1016,7	4,0	20,0	0,00	0,01
max. - min.	40,5	1017,0	12,0	20,0	0,13	
min. - max	40,0	1017,0	12,0	19,9	0,13	
random	38,3	1016,1	12,0	20,0	0,12	
Average	39,6	1016,7	12,0	20,0	0,13	0,01
max. - min.	80,5	1017,0	49,0	20,0	0,09	
min. - max	79,1	1017,0	49,0	19,9	0,11	
random	79,0	1016,1	48,0	20,0	0,11	
Average	79,5	1016,7	48,7	20,0	0,10	0,01
max. - min.	117,0	1017,0	101,0	20,0	0,11	
min. - max	118,0	1017,0	101,0	19,9	0,11	
random	118,0	1016,1	104,0	20,0	0,12	
Average	117,7	1016,7	102,0	20,0	0,11	0,01
max. - min.	160,0	1017,0	181,0	19,9	0,10	
min. - max	161,0	1017,0	181,0	19,9	0,11	
random	156,0	1016,1	179,0	20,0	0,10	
Average	159,0	1016,7	180,3	19,9	0,10	0,01
max. - min.	237,0	1017,0	414,0	19,9	0,22	
min. - max	237,0	1017,0	414,0	20,0	0,22	
random	242,0	1016,1	418,0	20,0	0,22	
Average	238,7	1016,7	415,3	20,0	0,22	0,00
max. - min.	311,0	1017,0	725,0	19,9	0,34	
min. - max	312,0	1017,0	725,0	19,9	0,35	
random	317,0	1016,1	725,0	20,0	0,37	
Average	313,3	1016,7	725,0	19,9	0,35	0,02
max. - min.	391,0	1017,0	1112,0	20,0	0,52	
min. - max	390,0	1017,0	1112,0	19,9	0,50	
random	390,0	1016,1	1135,0	20,0	0,52	
Average	390,3	1016,7	1119,7	20,0	0,51	0,01

Calibration result for Qmax. to Qmin., Qmin. to Qmax. and random

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter


6.1.2 Lab. 2. NMI

	Flow real m ³ /h	Pressure abs. at Pr mbar	Pressure drop. Pr-P Pa	Temperature °C	Error %	Standard diviation %
124	2,01	1046,0	1,0	20,6	-0,73	
125	2,01	1046,0	1,0	20,6	-0,71	
14	2,00	-	-	-	-0,58	
Average	2,01	1046,0	1,0	20,6	-0,67	0,08
46	5,1	1012,4	2,0	-	-0,17	
47	5,1	1012,4	2,0	-	-0,18	
48	5,1	1012,4	2,0	-	-0,17	
Average	5,1	1012,4	2,0	-	-0,17	0,01
40	10,2	1012,4	3,0	-	-0,04	
41	10,2	1012,4	3,0	-	-0,04	
42	10,2	1012,4	2,0	-	0,04	
Average	10,2	1012,4	2,7	-	-0,01	0,05
36	25,1	1012,4	8,0	-	0,07	
37	25,0	1012,4	8,0	-	0,06	
38	25,0	1012,4	8,0	-	0,09	
Average	25,0	1012,4	8,0	-	0,07	0,02
32	40,3	1012,4	16,0	-	0,09	
33	40,3	1012,4	16,0	-	0,09	
34	40,3	1012,4	16,0	-	0,14	
Average	40,3	1012,4	16,0	-	0,11	0,03
26	60,5	1012,4	32,0	-	0,14	
27	60,5	1012,4	32,0	-	0,14	
28	60,5	1012,4	32,0	-	0,12	
Average	60,5	1012,4	32,0	-	0,13	0,01
23	100,8	1012,4	80,0	-	0,18	
24	100,7	1012,4	80,0	-	0,17	
25	100,7	1012,4	80,0	-	0,18	
Average	100,7	1012,4	80,0	-	0,18	0,01
20	201,2	1011,8	292,0	-	0,25	
21	201,0	1011,8	292,0	-	0,27	
22	201,4	1011,8	294,0	-	0,26	
Average	201,2	1011,8	292,7	-	0,26	0,01
17	304,6	1011,8	667,0	-	0,39	
18	303,7	1011,8	662,0	-	0,38	
19	302,4	1011,8	656,0	-	0,39	
Average	303,6	1011,8	661,7	-	0,39	0,01
15	360,5	1011,8	945,0	-	0,47	
16	360,9	1011,8	945,0	-	0,47	
74	360,3	1011,8	946,0	-	0,48	
Average	360,6	1011,8	945,3	-	0,47	0,01
	-	-	-	-	-	
	-	-	-	-	-	
	-	-	-	-	-	
Average	-	-	-	-	-	-

Calibration in points 25, 60, 100, 200, 300 and 360 m³/h instead of 20, 80, 120, 160, 240 and 320 m³/h.

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.3 Lab. 3. Gaz de France, France

	Flow real m³/h	Pressure abs. at Pr mbar	Pressure drop. Pr-P PA	Temperature °C	Error %	Standard deviation %
	-	-	-	-	-	-
	-	-	-	-	-	-
	-	-	-	-	-	-
Average	-	-	-	-	-	-
	-	-	-	-	-	-
	-	-	-	-	-	-
	-	-	-	-	-	-
Average	-	-	-	-	-	-
14	10,08	1005	400	20,2	0,43	
15	10,08	1005	400	20,2	0,40	
16	10,08	1005	400	20,3	0,41	
Average	10,08	1005	400	2,0	0,41	0,02
55	20,42	1012	100	19,2	0,36	
54	20,44	1012	100	19,1	0,41	
53	20,44	1012	100	19,1	0,42	
Average	20,43	1012	100	19,1	0,40	0,03
61	40,42	1012	100	20,4	0,37	
60	40,43	1012	100	20,4	0,37	
59	40,47	1012	100	20,4	0,40	
Average	40,44	1012	100	20,4	0,38	0,02
21	79,97	1005	100	20,2	0,49	
22	80,26	1005	100	20,2	0,48	
23	80,45	1005	100	20,2	0,49	
Average	80,23	1005	100	20,2	0,49	0,01
17	121,1	1005	600	20,3	0,51	
19	121,2	1005	600	20,2	0,45	
18	121,2	1005	600	20,2	0,49	
Average	121,2	1005	600	20,2	0,48	0,03
65	158,5	1011	200	20,5	0,24	
66	158,7	1011	200	20,4	0,22	
67	158,8	1011	200	20,3	0,21	
Average	158,7	1011	200	20,4	0,22	0,02
51	237,9	1012	400	18,1	0,51	
52	237,9	1012	400	18,1	0,57	
50	238,1	1012	400	18,0	0,47	
Average	238,0	1012	400	18,1	0,52	0,05
64	317,0	1012	600	18,8	0,44	
63	317,3	1012	600	18,9	0,45	
62	317,4	1012	600	19,1	0,39	
Average	317,2	1012	600	18,9	0,43	0,03
56	395,9	1012	900	18,4	0,68	
57	396,9	1012	900	18,1	0,62	
58	398,7	1012	900	18,1	0,67	
Average	397,2	1012	900	18,2	0,66	0,03

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.4 Lab. 4. CESAME-LNE QUEST, France

	Flow real m³/h	Pressure abs. at Pr mbar	Pressure drop. Pr-P Pa	Temperature °C	Error %	Standard deviation %
-	-	-	-	-	-	-
-	-	-	-	-	-	-
-	-	-	-	-	-	-
Average	-	-	-	-	-	-
-	-	-	-	-	-	-
-	-	-	-	-	-	-
Average	-	-	-	-	-	-
max-min	10,079	993,80	0,0	20,7	-0,30	
random	10,057	995,90	-	20,6	-0,17	
random	10,057	996,00	-	20,6	-0,19	
Average	10,064	995,23	0,0	20,6	-0,22	0,07
max-min	20,200	993,70	0,0	20,6	-0,03	
random	20,179	993,90	-	20,6	-0,09	
random	20,179	994,00	-	20,7	-0,11	
Average	20,186	993,87	0,0	20,6	-0,08	0,04
max-min	39,867	993,80	10,0	20,5	0,07	
random	39,676	995,90	-	20,6	0,01	
random	39,676	996,00	-	20,6	-0,02	
Average	39,740	995,23	10,0	20,6	0,02	0,05
max-min	80,971	994,40	60,0	20,2	0,49	
random	80,971	994,80	-	20,3	0,46	
random	81,013	994,90	-	20,4	0,47	
Average	80,985	994,70	60,0	20,3	0,47	0,02
max-min	121,776	996,00	130,0	19,9	0,22	
random	123,693	995,20	-	20,4	0,23	
random	123,587	995,20	-	20,3	0,20	
Average	123,019	995,47	130,0	20,2	0,22	0,02
max-min	161,728	996,40	230,0	19,9	0,20	
random	164,072	996,10	-	20,8	0,23	
random	164,179	996,20	-	20,5	0,20	
Average	163,326	996,23	230,0	20,4	0,21	0,02
max-min	242,060	999,80	520,0	21,0	0,47	
random	245,469	999,20	-	20,3	0,46	
random	245,469	998,90	-	20,1	0,40	
Average	244,333	999,30	520,0	20,5	0,44	0,04
max-min	321,965	1003,40	930,0	19,5	0,21	
random	323,030	1003,00	-	20,4	0,25	
random	323,244	1003,00	-	19,9	0,32	
Average	322,746	1003,13	930,0	19,9	0,26	0,06
max-min	400,805	1009,60	1390,0	20,1	0,44	
random	393,560	1016,20	-	20,0	0,39	
random	394,199	1016,20	-	20,2	0,44	
Average	396,188	1014,00	1390,0	20,1	0,42	0,03

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.5 Lab. 5. Ministère des Affaires Economiques, Belgium

	Flow real m³/h	Pressure abs. at Pr mbar	Pressure drop. Pr-P Pa	Temperature °C	Error %	Standard deviation %
max. - min.	2,700	1023,87	1,0	22,0	-0,91	
min. - max	-	-	-	-	-	
random	2,040	1024,83	1,0	21,6	-0,99	
Average	2,370	1024,35	1,0	21,8	-0,95	0,06
max. - min.	5,350	1023,80	1,4	22,0	-0,39	
min. - max	6,200	1013,95	1,6	21,6	-0,41	
random	4,980	1019,94	1,5	21,8	-0,30	
Average	5,510	1019,23	1,5	21,8	-0,36	0,06
max. - min.	8,940	1023,67	1,8	22,1	-0,23	
min. - max	10,600	1023,26	2,0	21,7	-0,16	
random	10,010	1020,16	2,0	22,6	-0,11	
Average	9,850	1022,36	1,9	22,1	-0,17	0,06
max. - min.	20,050	1024,02	4,0	21,7	-0,12	
min. - max	19,830	1031,20	4,0	21,5	-0,10	
random	20,110	1024,00	4,0	21,4	-0,09	
Average	19,997	1026,41	4,0	21,5	-0,10	0,01
max. - min.	39,280	1026,04	12,0	21,9	-0,14	
min. - max	39,770	1031,40	13,0	21,3	-0,08	
random	39,420	1023,50	12,5	21,0	-0,15	
Average	39,490	1026,98	12,5	21,4	-0,12	0,04
max. - min.	79,650	1025,73	44,5	21,9	-0,11	
min. - max	80,370	1033,16	46,0	21,2	-0,01	
random	79,970	1023,69	46,0	21,2	-0,09	
Average	79,997	1027,53	45,5	21,5	-0,07	0,05
max. - min.	119,760	1024,92	97,0	21,9	-0,08	
min. - max	120,460	1032,82	98,0	21,1	0,01	
random	118,970	1019,70	95,0	22,5	-0,04	
Average	119,730	1025,81	96,7	21,8	-0,04	0,04
max. - min.	159,580	1024,92	163,0	21,9	-0,09	
min. - max	160,670	1032,80	166,0	21,1	0,00	
random	161,070	1023,21	165,0	21,7	0,03	
Average	160,440	1026,98	164,7	21,6	-0,02	0,06
max. - min.	240,010	1024,07	360,0	21,7	0,05	
min. - max	241,240	1032,14	365,0	21,0	0,12	
random	237,470	1022,19	356,0	20,9	-0,02	
Average	239,573	1026,13	360,3	21,2	0,05	0,07
max. - min.	319,760	1025,11	655,0	22,1	0,08	
min. - max	321,080	1033,19	665,0	21,0	0,18	
random	320,540	1020,85	655,0	20,7	0,14	
Average	320,460	1026,38	658,3	21,3	0,14	0,05
max. - min.	394,140	1023,87	1002,0	22,3	0,19	
min. - max	393,890	1031,88	1015,0	20,8	0,27	
random	393,280	1034,50	1010,0	21,2	0,27	
Average	393,770	1030,08	1009,0	21,4	0,24	0,05

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.6 Lab. 6. Eidg. Amt für Messwesen, Switzerland

	Flow real	Pressure abs. at Pr	Pressure drop. Pr-P	Temperature	Error	Standard diviation
	m³/h	mbar	Pa	°C	%	%
7	2,008	952,441	1,7	20,349	-0,85	
8	1,972	952,282	1,6	20,328	-0,93	
44	1,988	953,570	1,0	20,285	-0,93	
Average	1,989	952,764	1,4	20,321	-0,90	0,05
14	5,079	952,229	2,8	20,277	-0,25	
15	5,054	952,211	2,3	20,321	-0,21	
42	5,049	954,136	2,3	20,232	-0,21	
Average	5,061	952,859	2,5	20,277	-0,22	0,02
21	10,067	951,532	2,2	20,304	0,00	
22	10,070	951,406	2,1	20,326	-0,01	
23	10,073	951,381	2,2	20,348	-0,03	
Average	10,070	951,440	2,2	20,326	-0,01	0,02
28	20,290	950,411	3,3	20,269	0,05	
29	20,292	950,424	3,2	20,255	0,05	
45	20,152	955,112	3,8	19,994	0,05	
Average	20,245	951,982	3,4	20,173	0,05	0,00
33	40,738	956,212	15,0	20,193	0,08	
34	40,747	956,105	14,8	20,193	0,07	
35	40,751	956,014	14,9	20,208	0,06	
Average	40,745	956,110	14,9	20,198	0,07	0,01
39	80,611	954,617	49,3	20,167	0,11	
40	80,616	954,463	49,6	20,162	0,10	
41	80,614	954,355	49,4	20,162	0,11	
Average	80,614	954,478	49,4	20,164	0,11	0,01
36	119,953	955,863	104,8	20,165	0,14	
37	120,887	955,724	106,3	20,162	0,13	
38	120,885	955,647	106,8	20,162	0,14	
Average	120,575	955,745	106,0	20,163	0,14	0,01
30	159,850	956,374	181,4	20,219	0,14	
31	160,405	956,282	181,3	20,193	0,15	
32	160,415	956,269	181,5	20,178	0,15	
Average	160,223	956,308	181,4	20,197	0,15	0,01
24	240,162	950,933	393,8	20,247	0,25	
25	240,114	950,851	393,8	20,240	0,27	
26	240,085	950,802	393,9	20,224	0,27	
Average	240,120	950,862	393,8	20,237	0,26	0,01
17	320,111	951,898	711,8	20,109	0,39	
18	320,070	951,891	713,2	20,132	0,38	
19	320,184	951,867	712,6	20,131	0,40	
Average	320,122	951,885	712,5	20,124	0,39	0,01
2	398,962	953,730	1132,6	20,356	0,52	
3	399,118	953,748	1120,5	20,320	0,52	
4	398,568	953,597	1139,1	20,284	0,50	
Average	398,883	953,692	1130,7	20,320	0,51	0,01

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.7 Lab. 7. Bundesamt für Eich und Vermessungswesen, Austria

	Flow real	Pressure abs. at Pr	Pressure drop. Pr-P	Temperature	Error	Standard diviation
	m³/h	mbar	Pa	°C	%	%
max. - min.	2	990,7	0,0	21,7	-0,99	
min. - max	2	993,6	0,0	21,1	-0,99	
random	2	995,0	0,0	21,4	-0,98	
Average	2	993,1	0,0	21,4	-0,99	0,01
max. - min.	5	990,7	0,0	21,7	-0,33	
min. - max	5	993,6	0,0	21,1	-0,33	
random	5	995,0	0,0	21,3	-0,36	
Average	5	993,1	0,0	21,4	-0,34	0,02
max. - min.	10	990,7	10,0	21,7	-0,02	
min. - max	10	993,6	10,0	21,1	-0,03	
random	10	995,0	10,0	21,3	-0,03	
Average	10	993,1	10,0	21,4	-0,03	0,01
max. - min.	20	990,7	10,0	21,6	0,12	
min. - max	20	993,6	10,0	21,1	0,12	
random	20	995,0	10,0	21,2	0,10	
Average	20	993,1	10,0	21,3	0,11	0,01
max. - min.	40	990,7	20,0	21,6	0,13	
min. - max	40	993,6	20,0	21,2	0,15	
random	40	995,0	20,0	21,1	0,13	
Average	40	993,1	20,0	21,3	0,14	0,01
max. - min.	80	990,7	50,0	21,5	0,07	
min. - max	80	993,6	50,0	21,2	0,06	
random	80	995,0	50,0	21,0	0,00	
Average	80	993,1	50,0	21,2	0,04	0,04
max. - min.	120	990,7	110,0	21,5	0,03	
min. - max	120	993,6	110,0	21,2	0,07	
random	120	995,0	110,0	21,0	0,07	
Average	120	993,1	110,0	21,2	0,06	0,02
max. - min.	160	990,7	180,0	21,3	0,16	
min. - max	160	993,6	180,0	21,3	0,17	
random	160	995,0	180,0	20,9	0,16	
Average	160	993,1	180,0	21,2	0,16	0,01
max. - min.	240	990,7	410,0	21,2	0,34	
min. - max	240	993,6	410,0	21,4	0,35	
random	240	995,0	410,0	20,8	0,32	
Average	240	993,1	410,0	21,1	0,34	0,02
max. - min.	320	990,7	720,0	21,2	0,40	
min. - max	320	993,6	720,0	21,4	0,40	
random	320	995,0	720,0	20,7	0,40	
Average	320	993,1	720,0	21,1	0,40	0,00
max. - min.	400	990,7	1140,0	21,1	0,58	
min. - max	400	993,6	1140,0	21,4	0,59	
random	400	995,0	1140,0	20,6	0,57	
Average	400	993,1	1140,0	21,0	0,58	0,01

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.8 Lab. 8. Instiuto de Metrologia "G. Colonnetti", Italy

	Flow real m³/h	Pressure abs. at Pr mbar	Pressure drop. Pr-P Pa	Temperature °C	Error %	Standard deviation %
max. - min.	2,090	996,16	1,4	19,9	-0,74	
min. - max	2,080	996,14	1,0	19,9	-0,75	
random	2,090	995,91	1,3	20,1	-0,72	
Average	2,087	996,07	1,2	20,0	-0,74	0,02
max. - min.	5,000	996,14	1,1	19,8	-0,21	
min. - max	5,000	996,10	1,4	20,0	-0,23	
random	5,000	994,41	5,0	20,3	-0,22	
Average	5,000	995,55	2,5	20,0	-0,22	0,01
max. - min.	9,900	996,14	2,0	19,8	-0,02	
min. - max	9,890	996,09	2,0	20,0	-0,04	
random	9,910	996,03	2,0	20,0	-0,02	
Average	9,900	996,09	2,0	19,9	-0,03	0,01
max. - min.	20,080	996,15	4,7	19,8	0,10	
min. - max	20,080	996,07	4,6	20,0	0,10	
random	20,090	993,98	1,4	20,3	0,08	
Average	20,083	995,40	3,6	20,0	0,09	0,01
max. - min.	38,980	996,20	13,8	19,7	0,09	
min. - max	39,000	996,07	13,4	20,0	0,08	
random	39,020	995,81	13,6	20,1	0,07	
Average	39,000	996,02	13,6	19,9	0,08	0,01
max. - min.	80,110	996,25	50,0	19,7	0,13	
min. - max	80,190	996,75	51,1	20,0	0,11	
random	80,240	996,01	51,0	20,1	0,10	
Average	80,180	996,34	50,7	19,9	0,11	0,02
max. - min.	120,300	996,28	109,1	19,7	0,14	
min. - max	120,490	996,08	104,8	20,0	0,10	
random	120,550	995,73	106,5	20,1	0,11	
Average	120,447	996,03	106,8	19,9	0,12	0,02
max. - min.	160,600	996,30	185,8	19,6	0,19	
min. - max	160,940	996,07	185,8	20,0	0,13	
random	161,060	994,41	190,5	20,3	0,09	
Average	160,867	995,59	187,4	20,0	0,14	0,05
max. - min.	-	-	-	-	-	-
min. - max	-	-	-	-	-	-
random	-	-	-	-	-	-
Average	-	-	-	-	-	-
max. - min.	-	-	-	-	-	-
min. - max	-	-	-	-	-	-
random	-	-	-	-	-	-
Average	-	-	-	-	-	-
max. - min.	-	-	-	-	-	-
min. - max	-	-	-	-	-	-
random	-	-	-	-	-	-
Average	-	-	-	-	-	-

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.9 Lab. 9. Országos Mérésügyi Hivatal, Hungary

	Flow real m³/h	Pressure abs. at Pr mbar	Pressure drop. Pr-P Pa	Temperature °C	Error %	Standard deviation %
12	2,360	995,00	0,0	21,3	-1,23	
1	1,990	994,80	0,0	21,4	-1,43	
1	2,110	995,50	0,0	21,2	-1,02	
Average	2,153	995,10	0,0	21,3	-1,23	0,21
11	4,760	994,90	0,0	21,2	-0,51	
2	4,850	994,90	0,0	21,3	-0,31	
2	4,820	994,30	0,0	21,0	-0,41	
Average	4,810	994,70	0,0	21,2	-0,41	0,10
10	8,080	994,50	3,0	21,2	-0,15	
3	8,110	995,20	3,0	21,2	0,05	
3	7,910	994,90	3,0	21,1	-0,10	
Average	8,033	994,87	3,0	21,2	-0,07	0,10
8	20,040	994,70	5,0	21,1	-0,20	
5	19,800	994,90	5,0	21,2	-0,14	
5	20,080	993,30	5,0	21,0	-0,15	
Average	19,973	994,30	5,0	21,1	-0,16	0,03
7	39,800	994,90	15,0	21,0	-0,22	
6	38,360	994,80	15,0	21,1	-0,23	
6	38,670	995,40	15,0	21,0	-0,25	
Average	38,943	995,03	15,0	21,0	-0,23	0,02
6	79,720	994,20	50,0	20,9	-0,14	
7	79,070	994,50	50,0	20,9	-0,14	
7	78,300	992,70	50,0	21,0	-0,16	
Average	79,030	993,80	50,0	20,9	-0,15	0,01
5	120,120	994,50	110,0	20,9	-0,02	
8	120,900	994,40	110,0	20,9	-0,02	
8	114,980	995,80	110,0	20,9	-0,04	
Average	118,667	994,90	110,0	20,9	-0,03	0,01
4	161,340	994,10	190,0	20,9	0,02	
9	158,080	994,20	190,0	20,9	0,01	
9	157,460	995,70	190,0	20,9	0,03	
Average	158,960	994,67	190,0	20,9	0,02	0,01
3	236,870	993,40	410,0	20,9	0,08	
10	233,460	993,70	410,0	20,8	0,10	
10	231,710	995,80	410,0	20,9	0,09	
Average	234,013	994,30	410,0	20,9	0,09	0,01
2	320,170	992,30	710,0	20,9	0,07	
11	308,080	993,10	710,0	20,8	0,05	
11	308,170	996,20	710,0	20,8	0,05	
Average	312,140	993,87	710,0	20,8	0,06	0,01
1	399,660	991,30	1220,0	20,8	0,26	
12	396,120	991,90	1220,0	20,9	0,25	
12	390,590	996,30	1220,0	20,8	0,20	
Average	395,457	993,17	1220,0	20,8	0,24	0,03

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.10 Lab. 10. Czech Metrological Institute, Czech Rep.

	Flow real m ³ /h	Pressure abs. at Pr mbar	Pressure drop. Pr-P Pa	Temperature °C	Error %	Standard deviation %
-	1,980	1003,41	0,0	21,1	-0,85	
-	1,980	1003,22	0,0	21,2	-0,88	
-	1,980	1003,41	0,0	21,3	-0,88	
Average	1,980	1003,35	0,0	21,2	-0,87	0,02
-	5,030	1003,39	0,0	21,1	-0,18	
-	5,030	1003,22	0,0	21,2	-0,18	
-	5,030	1003,53	0,0	21,3	-0,12	
Average	5,030	1003,38	0,0	21,2	-0,16	0,03
-	10,030	1003,42	0,0	21,1	-0,12	
-	10,020	1003,25	0,0	21,1	-0,13	
-	10,020	1003,33	0,0	21,3	-0,11	
Average	10,023	1003,33	0,0	21,2	-0,12	0,01
-	20,050	1003,42	0,0	21,1	-0,04	
-	20,390	1003,21	0,0	21,1	-0,06	
-	20,420	1003,37	0,0	21,3	-0,03	
Average	20,287	1003,33	0,0	21,2	-0,04	0,02
-	39,940	1003,41	2,0	21,1	-0,01	
-	39,920	1003,22	2,0	21,1	-0,02	
-	39,950	1003,43	2,0	21,5	0,01	
Average	39,937	1003,35	2,0	21,2	-0,01	0,02
-	80,250	1003,31	30,0	21,1	0,02	
-	80,250	1003,13	31,0	21,2	0,03	
-	80,260	1003,21	30,0	21,3	0,01	
Average	80,253	1003,22	30,3	21,2	0,02	0,01
-	120,290	1003,17	90,0	21,1	0,07	
-	120,260	1003,04	92,0	21,2	0,08	
-	120,290	1003,07	91,0	21,3	0,07	
Average	120,280	1003,09	91,0	21,2	0,07	0,01
-	160,150	1003,02	165,0	21,1	0,05	
-	160,130	1002,87	164,0	21,2	0,05	
-	160,160	1003,05	167,0	21,3	0,04	
Average	160,147	1002,98	165,3	21,2	0,05	0,01
-	239,750	1002,79	378,0	21,1	0,20	
-	239,720	1002,79	391,0	21,2	0,20	
-	239,690	1002,94	384,0	21,3	0,19	
Average	239,720	1002,84	384,3	21,2	0,20	0,01
-	318,150	1002,06	693,0	21,1	0,30	
-	318,020	1001,86	706,0	21,1	0,32	
-	318,010	1001,96	690,0	21,1	0,31	
Average	318,060	1001,96	696,3	21,1	0,31	0,01
-	395,320	1001,12	1104,0	21,0	0,48	
-	395,440	1000,93	1117,0	21,3	0,45	
-	395,640	1001,16	1107,0	21,4	0,46	
Average	395,467	1001,07	1109,3	21,2	0,46	0,02

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.11 Lab. 11. PTB, Germany

	Flow real m³/h	Pressure abs. at Pr mbar	Pressure drop. Pr-P Pa	Temperature °C	Error %	Standard deviation %
11	2,122	1003,76	1,0	20,5	1,26	
12	2,105	1003,71	1,0	20,5	0,44	
23	2,129	1003,80	1,0	20,5	1,63	
Average	2,119	1003,76	1,0	20,5	1,11	0,61
10	5,217	1003,74	1,0	20,6	-0,26	
13	5,216	1003,74	1,0	20,5	-0,26	
26	5,216	1003,90	2,0	20,5	-0,26	
Average	5,216	1003,79	1,3	20,5	-0,26	0,00
9	10,307	1003,74	2,0	20,6	-0,07	
14	10,306	1003,69	2,0	20,6	-0,08	
28	10,304	1003,87	2,0	20,4	-0,07	
Average	10,306	1003,77	2,0	20,5	-0,07	0,00
8	20,596	1003,75	5,0	20,6	0,02	
15	20,592	1003,77	5,0	20,6	0,01	
30	20,592	1003,86	5,0	20,4	0,02	
Average	20,593	1003,79	5,0	20,5	0,02	0,00
7	41,160	1003,73	15,0	20,6	0,09	
16	41,155	1003,75	15,0	20,6	0,08	
32	41,152	1003,86	15,0	20,3	0,08	
Average	41,156	1003,78	15,0	20,5	0,08	0,00
6	82,224	1003,69	52,0	20,6	0,10	
17	82,219	1003,62	52,0	20,6	0,10	
25	82,202	1003,70	52,0	20,5	0,09	
Average	82,215	1003,67	52,0	20,6	0,10	0,00
5	124,016	1003,54	113,0	20,7	0,15	
18	123,967	1003,47	113,0	20,5	0,14	
33	123,916	1003,60	112,0	20,3	0,15	
Average	123,966	1003,54	112,7	20,5	0,15	0,00
4	163,125	1003,31	190,0	20,7	0,18	
19	163,060	1003,25	191,0	20,5	0,18	
31	163,020	1003,33	191,0	20,4	0,18	
Average	163,068	1003,30	190,7	20,5	0,18	0,00
3	245,590	1002,83	422,0	20,7	0,29	
20	245,442	1002,60	422,0	20,4	0,29	
24	245,491	1002,70	423,0	20,5	0,29	
Average	245,508	1002,71	422,3	20,5	0,29	0,00
2	326,251	1001,96	751,0	20,7	0,41	
21	326,070	1001,72	749,0	20,4	0,41	
29	326,006	1003,78	750,0	20,3	0,41	
Average	326,109	1002,49	750,0	20,5	0,41	0,00
1	406,928	1000,79	1193,0	20,6	0,60	
22	406,846	1000,61	1188,0	20,4	0,60	
27	406,650	1000,60	1189,0	20,3	0,60	
Average	406,808	1000,67	1190,0	20,4	0,60	0,00

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.12 Lab. 12. Slovak Institute of Metrology, Slovakian

	Flow real	Pressure abs. at Pr	Pressure drop. Pr-P	Temperature	Error	Standard deviation
	m³/h	mbar	Pa	°C	%	%
31-33	1,970	989,90	0,0	22,6	-0,64	
34-36	1,970	989,30	0,0	22,6	-0,65	
Average	1,970	989,60	0,0	22,6	-0,65	0,01
28-30	5,191	990,40	0,0	22,6	-0,20	
37-39	5,192	989,30	0,0	22,6	-0,23	
Average	5,192	989,85	0,0	22,6	-0,22	0,02
25-27	10,170	991,50	5,0	22,6	-0,12	
40-42	10,170	989,20	5,0	22,6	-0,13	
Average	10,170	990,35	5,0	22,6	-0,13	0,01
22-24	20,080	991,65	10,0	22,6	-0,02	
43-45	20,080	989,15	10,0	22,6	-0,03	
Average	20,080	990,40	10,0	22,6	-0,03	0,01
19-21	40,060	988,90	20,0	21,8	0,05	
46-48	40,060	987,60	20,0	22,2	0,03	
Average	40,060	988,25	20,0	22,0	0,04	0,01
16-18	80,270	989,00	70,0	22,0	0,08	
49-51	80,270	987,40	60,0	22,2	0,06	
Average	80,270	988,20	65,0	22,1	0,07	0,01
13-15	120,400	988,80	140,0	21,7	0,07	
52-54	120,400	987,10	140,0	22,2	0,08	
Average	120,400	987,95	140,0	22,0	0,08	0,01
10-12	160,500	988,70	250,0	21,9	0,09	
55-57	160,600	986,90	250,0	22,2	0,11	
Average	160,550	987,80	250,0	22,0	0,10	0,01
7-9	240,800	988,10	550,0	22,0	0,22	
58-60	240,900	986,40	560,0	22,1	0,21	
Average	240,850	987,25	555,0	22,0	0,22	0,01
4-6	321,000	987,40	980,0	22,1	0,35	
61-63	321,100	985,60	990,0	22,1	0,34	
Average	321,050	986,50	985,0	22,1	0,35	0,01
1-3	401,600	986,50	1500,0	22,0	0,49	
64-66	401,800	984,60	1530,0	22,1	0,48	
Average	401,700	985,55	1515,0	22,0	0,49	0,01

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

6.1.13 Lab. 13. Central Office of Measures, Poland


Central Office of Measures

Lab. 13

	Flow real m³/h	Pressure abs. at Pr mbar	Pressure drop. Pr-P Pa	Temperature °C	Error %	Standard deviation %	U %
-	-	-	-	-	-	-	-
-	-	-	-	-	-	-	-
-	-	-	-	-	-	-	-
Average	-	-	-	-	-	-	-
-	-	-	-	-	-	-	-
-	-	-	-	-	-	-	-
-	-	-	-	-	-	-	-
Average	-	-	-	-	-	-	-
-	-	-	-	-	-	-	-
-	-	-	-	-	-	-	-
-	-	-	-	-	-	-	-
Average	-	-	-	-	-	-	-
8	20,040	1057,60	-	20,9	0,17		
1	20,130	1057,60	-	20,8	0,18		
8	20,070	1071,40	-	21,0	0,31		
Average	20,080	1062,20	-	20,9	0,22	0,08	0,26
7	41,150	1057,50	-	20,9	0,00		
2	39,130	1057,50	-	20,8	0,01		
7	39,260	1071,35	-	21,0	0,01		
Average	39,847	1062,12	-	20,9	0,01	0,01	0,24
6	79,290	1057,40	-	20,9	0,02		
3	80,080	1057,40	-	20,8	-0,04		
6	78,600	1071,20	-	21,0	0,01		
Average	79,323	1062,00	-	20,9	0,00	0,03	0,22
5	118,210	1057,25	-	20,9	-0,01		
4	118,810	1057,30	-	20,7	-0,02		
5	118,530	1071,15	-	21,0	-0,01		
Average	118,517	1061,90	-	20,9	-0,01	0,01	0,22
4	161,050	1057,10	-	20,9	-0,02		
5	157,180	1057,10	-	20,7	-0,01		
4	157,250	1070,90	-	21,0	-0,06		
Average	158,493	1061,70	-	20,9	-0,03	0,03	0,22
3	238,750	1056,50	-	20,9	0,01		
6	236,330	1056,60	-	20,7	0,01		
3	233,670	1070,40	-	21,0	-0,03		
Average	236,250	1061,17	392	20,9	0,00	0,02	0,22
2	323,130	1056,00	-	20,9	0,15		
7	318,340	1055,80	-	20,8	0,03		
2	317,360	1069,60	-	21,0	0,04		
Average	319,610	1060,47	-	20,9	0,07	0,07	0,22
1	404,460	1054,80	-	20,9	0,24		
8	405,010	1054,80	-	20,8	0,20		
1	402,870	1068,60	-	21,0	0,21		
Average	404,113	1059,40	-	20,9	0,22	0,02	0,22

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter

6.1.14 Lab. 14. NEL, United Kingdom


NEL

Lab. 14

	Flow real	Pressure abs. at Pr	Pressure drop. Pr-P	Temperature	Error	Standard deviation	U
	m ³ /h	mbar	Pa	°C	%	%	%
33	2,110	-	1,0	21,6	-0,02		
35	2,100	-	1,0	21,4	-0,02		
36	2,100	-	1,0	21,3	-0,09		
Average	2,103	-	1,0	21,4	-0,04	0,04	0,37
31	5,060	-	1,0	22,1	0,53		
32	5,060	-	1,0	21,9	0,36		
34	5,060	-	1,0	21,5	0,37		
Average	5,060	-	1,0	21,8	0,42	0,10	0,37
1	10,060	-	1,0	21,2	0,27		
2	10,070	-	1,0	20,7	0,39		
3	10,050	-	1,0	20,1	0,30		
Average	10,060	-	1,0	20,7	0,32	0,06	0,37
4	19,940	-	6,0	21,8	0,11		
6	19,990	-	6,0	21,2	0,14		
8	19,910	-	6,0	21,4	0,10		
Average	19,947	-	6,0	21,5	0,12	0,02	0,37
5	39,300	-	16,0	21,6	-0,01		
7	39,250	-	16,0	21,0	0,04		
9	39,220	-	16,0	21,4	0,03		
Average	39,257	-	16,0	21,3	0,02	0,03	0,37
11	79,050	-	59,0	19,1	0,18		
14	78,990	-	60,0	19,0	0,24		
15	79,060	-	60,0	19,0	0,22		
Average	79,033	-	59,7	19,0	0,21	0,03	0,37
10	118,270	-	126,0	19,1	0,19		
12	118,110	-	127,0	18,7	0,16		
13	118,150	-	127,0	19,0	0,20		
Average	118,177	-	126,7	18,9	0,18	0,02	0,37
16	158,680	-	219,0	18,2	0,26		
23	159,670	-	220,0	19,7	0,30		
27	160,280	-	218,0	20,1	0,27		
Average	159,543	-	219,0	19,3	0,28	0,02	0,37
17	240,570	-	500,0	17,7	0,35		
22	241,700	-	501,0	19,8	0,39		
26	242,380	-	500,0	20,1	0,37		
Average	241,550	-	500,3	19,2	0,37	0,02	0,37
18	323,460	-	890,0	21,5	0,49		
21	322,420	-	889,0	20,2	0,50		
25	323,350	-	891,0	20,7	0,47		
Average	323,077	-	890,0	20,8	0,49	0,02	0,37
19	400,980	-	1402,0	19,6	0,62		
20	400,550	-	1404,0	18,6	0,63		
30	401,150	-	1403,0	19,4	0,64		
Average	400,893	-	1403,0	19,2	0,63	0,01	0,37

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Euromet Project No. 419


Intercomparison of calibrations of a G250 rotary-meter

6.1.15 All the laboratories

Flow nominal	Error %													
	Lab1	Lab2	Lab3	Lab4	Lab5	Lab6	Lab7	Lab8	Lab9	Lab10	Lab11	Lab12	Lab13	Lab14
2	-0.41	-0.67			-0.95	-0.90	-0.99	-0.74	-1.23	-0.87	1.11	-0.65		-0.04
5	-0.06	-0.17			-0.36	-0.22	-0.34	-0.22	-0.41	-0.16	-0.26	-0.22		0.42
10	-0.04	-0.01	0.41	-0.22	-0.17	-0.01	-0.03	-0.03	-0.07	-0.12	-0.07	-0.13		0.32
20	0.00	0.07	0.40	-0.08	-0.10	0.05	0.11	0.09	-0.16	-0.04	0.02	-0.03	0.22	0.12
40	0.13	0.11	0.38	0.02	-0.12	0.07	0.14	0.08	-0.23	-0.01	0.08	0.04	0.01	0.02
80	0.10	0.13	0.49	0.47	-0.07	0.11	0.04	0.11	-0.15	0.02	0.10	0.07	0.00	0.21
120	0.11	0.18	0.48	0.22	-0.04	0.14	0.06	0.12	-0.03	0.07	0.15	0.08	-0.01	0.18
160	0.10	0.26	0.22	0.21	-0.02	0.15	0.16	0.14	0.02	0.05	0.18	0.10	-0.03	0.28
240	0.22	0.39	0.52	0.44	0.05	0.26	0.34		0.09	0.20	0.29	0.22	0.00	0.37
320	0.35	0.47	0.43	0.26	0.14	0.39	0.40		0.06	0.31	0.41	0.35	0.07	0.49
400	0.51		0.66	0.42	0.24	0.51	0.58		0.24	0.46	0.60	0.49	0.22	0.63

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter




6.1.16 Average error, Median erro and weighing error

Flow m³/h.	Average		
	Error %	Median %	weighing average %
2	-0.58	-0.74	-0.43
5	-0.18	-0.22	-0.21
10	-0.01	-0.04	-0.03
20	0.05	0.03	0.06
40	0.05	0.06	0.07
79	0.12	0.10	0.11
119	0.12	0.12	0.13
163	0.13	0.14	0.14
245	0.26	0.26	0.27
323	0.32	0.35	0.34
398	0.46	0.50	0.48

Euromet Project No. 419

Intercomparison of calibrations of a G250 rotary-meter



Jesper Busk
FORCE Technology